



NAME B26412
 EXPNO 400
 PROCNO 1
 Date_ 20190930
 Time_ 11.47
 INSTRUM spect
 PROBHD 5mm dual

PULPROG zg30
 TD 65536
 SOLVENT DMSO
 NS 8
 DS 0
 SWH 8620.689 F
 FIDRES 0.131541 F
 AQ 3.8011379 s
 RG 203
 DW 58.000 s
 DE 6.50 s
 TE 300.0 F
 D1 1.00000000 s
 TD0 1

===== CHANNEL f1 =====
 NUC1 1H
 P1 13.22 s
 PL1 -3.00 c
 PL1W 20.99974442 V
 SFO1 400.1319006 M
 SI 32768
 SF 400.1300061 M
 WDW EM
 SSB 0
 LB 0.30 F
 GB 0
 PC 1.00