

MASS SPECTROMETRY REPORT



| | | | | | |
|--------------------|-------------------|---------------------|----------|---------------|---------------|
| Sample Description | | Instrument | | Waters ZQ2000 | |
| Analyzed date: | 2018-07-13 | Probe: | ESI | Probe Bias: | +4.5kv |
| Analyst: | PY | Nebulizer Gas Flow: | 1.5L/min | Detector: | 1.5kv |
| Sample: | [Leu5]-Enkephalin | CDL: | -20.0v | T. Flow: | 0.2ml/min |
| M.W.: | 555.62 | CDL Temp.: | 250 °C | B. Conc.: | 50%H2O/50%ACN |
| Lot. No.: | P180627-1 | Block Temp.: | 200 °C | | |