



Current Data Parameters
NAME 2019-03-21 19W157B-hetengzigan
EXPNO 1
PROCNO 1

F2 - Acquisition Parameters
Date_ 20190321
Time_ 14.44 h
INSTRUM spect
PROBHD zg30
PULPROG zg30
TD 65536
SOLVENT MeOD
NS 16
DS 0
SWH 12019.230 Hz
FIDRES 0.366798 Hz
AQ 2.7262976 sec
RG 81.39
DW 41.600 usec
DE 6.50 usec
TE 298.2 K
D1 1.00000000 sec
TD0 1
SFO1 600.1137057 MHz
NUC1 1H
P1 15.00 usec
PLW1 19.4699931 W

F2 - Processing parameters
SI 65536
SF 600.1100000 MHz
WDW GM
SSB 0
LB -0.10 Hz
GB 0.1
PC 1.00

